

Certificate of Calibration

Issued By

CoMech Metrology Limited

Date of issue

29-Oct-2024

Certificate number

C458442-01

CoMech®

Calibration House
Unit 1 Victoria Way, Pride Park
Derby DE24 8AN

Tel: 01332 867700

Fax: 01332 867707

Email: metrology@comech.co.uk

Page 1 of 1

Authorised Signatory:
Ashley Lenton



Client : Catena Inspection Engineering Services Ltd
Brookhill Road, Brookhill Road Industrial Estate, Pinxton, Notts, NG16 6NS, UK

Order No : CASH

Manufacturer / Model No : Testo 540

Description : Light Meter

Identification : 1902709 **Serial No :** 39057744/505

Date received : 29-Oct-2024 **Date of test :** 29-Oct-2024

Basis of test : Examined and measured using a Light Source and a Light Meter at an ambient temperature of 20°C +/- 5°C to the manufacturer's specification and the measured values reported.

Test Report

| <u>Input</u> | <u>Indicated</u> | <u>Spec</u> | <u>Error</u> |
|--------------|------------------|-------------|--------------|
| | | +/- | |
| 420 Lux | 431 Lux | 13 | -11 |
| 1983 Lux | 2006 Lux | 60 | -23 |
| 28200 Lux | 27790 Lux | 834 | 410 |
| 85600 Lux | 86760 Lux | 2603 | -1160 |

Comments:

All equipment used has calibration certification traceable to National Standards.

Asset conforms to manufacturers specifications.

Equipment used: **HP453 & HP601**

This certificate may not be reproduced other than in full, except with prior written approval of the issuing laboratory.