

Certification of Calibration



Issued By	Staht Connect	Address	Unit 10 Hayes Lane Trading Estate, Folkes Road, Stourbridge, DY9 8RG
Telephone	0121 817 0620	600-044	25kN t25 Digital Pull Tester (Case)
Serial No	25-118	Certificate No	STC0002079

CALIBRATED REFERENCE READING	2.5	5	10	15	20	25
PRODUCT OUTPUT READING	2.5	5.0	10.1	15.0	19.9	24.9

All readings are in kN. All readings are within $\pm 1.5\%$ F.S.D.

The Staht 25kN t25 Digital Tensile Tester has been calibrated using equipment traceable to UKAS (certificate no. 15393). This calibration has been performed using the manufacturer's instructions that are traceable to National Standards.

Temperature : 21.2° C

Approved Signatory Robert Hirst

Robert Hirst
Managing Director

Calibration Date 05/11/2025

Calibration Expiry Date 05/11/2026

