

# Certificate of Calibration

Issued By

CoMech Metrology Limited

Date of issue

31-Oct-2025

Certificate number

C493635-01

Calibrated at

**CoMech®**

Calibration House  
1 Riverside Court, Pride Park,  
Derby, DE24 8JN

Page 1 of 1

Tel: 01332 867700

Email: metrology@comech.co.uk



**Client :** Catena Inspection Engineering Services Ltd  
Brookhill Road, Brookhill Road Industrial Estate, Pinxton, Notts, NG16 6NS, UK

**Order No :** PO242499

**Manufacturer / Model No :** Testo 540

**Description :** Light Meter

**Identification :** 1902709      **Serial No :** 39057744/505

**Date received :** 24-Oct-2025      **Date of test :** 31-Oct-2025

**Basis of test :** Examined and measured using a Light Source and a Light Meter at an ambient temperature of 20°C +/- 5°C unless otherwise stated, the equipment conforms in all respects to manufacturers published specifications at the points measured.

## Test Report

	<u>Input</u>	<u>Indicated</u>	<u>Spec</u>	<u>Error</u>
<i>RANGE</i>			+/-	
40 LUX	42.5 Lux	43.00 Lux	4.29 Lux	-0.50 Lux
400 LUX	384.8 Lux	398.0 Lux	14.9 Lux	-13.2 Lux
4000 LUX	3891 Lux	3861 Lux	119 Lux	30 Lux
40k LUX	39.18 k Lux	39.42 k Lux	4.18 k Lux	-0.24 k Lux

### Comments:

All equipment used has calibration certification traceable to National Standards.

Asset conforms to manufacturers specification.

Equipment used: **HP601 & HP453**

This certificate may not be reproduced other than in full, except with prior written approval of the issuing laboratory.