

**Dynamic Load Monitoring (UK) Ltd**

DLM House, Bridgers Farm, Nursling Street, Nursling, Southampton, SO16 0YA

T: +44 (0)2380 741 700 E: sales@dml-uk.com W: dlm-uk.com

CERTIFICATE OF CALIBRATION IN COMPRESSION

Customer:	Catena Inspection & Engineering Services Ltd	Certificate number:	DLM 0026 00323
Calibrated on:	03/08/2026	Recalibration due:	02/08/2027
Customer code:	CATENA	Customer order:	PO261902
Sales order:	52707/1	Customer Reference:	

Loadcell details

Manufacturer:	Dynamic Load Monitoring (UK) Ltd		
Product type:	Compressive loadcell + Handheld display		
Serial number:	DLM11796 & 10301412	Plant / Inventory no:	
Sensing element:	Strain Bridge	Resistance(Ω):	1000
Safe working load:	100 t	Sensitivity at SWL:	0.638 mV/V
Proof load:	150% S.W.L	Average error:	0.00 %

Test machine details

MachineType:	SANS SHT 4106G 100t Test Machine	Serial number:	31307019
Calibrated to:	BS EN ISO 7500-1:2018 Class 0.5 1000kN-2kN	Certificate number:	26-06167 (UKAS)

Readings (t) Test machine	Loadcell readings (Tonnes)				Total errors
	Run 1	Run 2	Run 3		
0.00	0.000	0.000	0.000		0.00 %
20.00	20.000	20.000	20.000		0.00 %
40.00	40.000	40.000	40.000		0.00 %
60.00	60.000	60.000	60.010		0.01 %
80.00	80.000	80.000	80.010		0.00 %
100.00	100.000	100.000	100.010		0.00 %

The uncertainties are for a confidence probability of not less than 95%

Certifying engineer:

Seb Nowakowski

VAT REG NO: GB 568 4862 87

REGISTERED IN ENGLAND NO. 2924110

REGISTERED OFFICE DLM HOUSE, BRIDGERS FARM, NURSING STREET, SOUTHAMPTON, HAMPSHIRE, SO16 0YA